

Quality Control

Certificate of Analysis

QC release date: 2021-04-13

Product: Alkaline phosphatase FS IFCC 37 °C

Product Code: 10441

Cat. No.: 1 0441 99 10 021

Delivery of kits: No of kits: 120
Kit Lot/Los: 60145249
Kit Expiry: 2022-07-01

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 30783
Expiry: 2022-07-01

Reagent 2 Lot No.: 30783
Expiry: 2022-07-01

Examination:

Appearance: Reagent 1: clear, colourless
Reagent 2: clear, yellowish

Odour: Reagent 1: slight smell (amines)
Reagent 2: no characteristic odour

pH: Reagent 1 (30 °C ± 1 °C): 10.37
Reagent 2 (25 °C ± 1 °C): 8.93
R1+R2 (30 °C ± 1 °C): 10.38

Absorbance: Reagent 1: 0.000
405 nm / 25 °C R1 + R2: 0.457

Control sera: TruLab N TruLab P
Lot: 25894 Lot: 25895

Fresh (at 37°C) 71.1 U/L 156 U/L

Target values 74.6 56.0 - 93.3 167 125 - 209

Additional remarks: Mf-Date: 01.04.2021

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka