

Quality Control

Certificate of Analysis

QC release date: 2021-04-08

Product: Apolipoprotein A1 FS
Product Code: 17102

Cat. No.: 1 7102 99 10 021

Delivery of kits: No of kits: 30
Kit Lot/Los: 60149022
Kit Expiry: 2023-03-26

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 30700
Expiry: 2023-03-26

Reagent 2 Lot No.: 30700
Expiry: 2023-03-26

Examination:

Appearance: Reagent 1: clear to slightly opalescent
Reagent 2: clear, slightly brownish

Odour: Reagent 1: no characteristic odour
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 7.47
Reagent 2 (25 °C ± 1 °C): 7.45

Absorbance: Reagent 1: 0.001
580 nm / 37 °C R1 + R2: 0.005

Control sera: TruLab L 1 TruLab L 2
Lot: 27037 Lot: 27038

Test results 104 mg/dL 200 mg/dL

Target values 106 84.8 - 127 205 164 - 246

Additional remarks: Mf-Date: 26.03.2021

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed
by hand.

Sign.: Dr. G. Gorka