

Quality Control

## Certificate of Analysis

QC release date: 2020-09-28

**Product:** Apolipoprotein B FS

**Product Code:** 17112

Cat. No.: 1 7112 99 10 021

---

**Delivery of kits:** No of kits: 1  
Kit Lot/Los: 60142341  
Kit Expiry: 2022-09-11

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 29091  
Expiry: 2022-09-11

Reagent 2 Lot No.: 29091  
Expiry: 2022-09-11

### Examination:

Appearance: Reagent 1: clear to slightly opalescent

Reagent 2: clear, brownish

Odour: Reagent 1: no characteristic odour  
Reagent 2: no characteristic odour

**pH:** Reagent 1 ( 25 °C ± 1 °C): 7.49  
Reagent 2 ( 25 °C ± 1 °C): 7.43

**Absorbance:** Reagent 1: 0.034  
340 nm / 37 °C R1 + R2: 0.081

**Control sera:** TruLab L 1 TruLab L 2  
Lot: 27037 Lot: 27038

Test results 82.7 mg/dL 133 mg/dL

Target values 84.1 67.2 - 101 134 107 - 160

---

**Additional remarks:** Mf-Date: 11.09.2020

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed  
by hand.

Sign.: Dr. G. Gorka