

Quality Control

## Certificate of Analysis

QC release date: 2021-08-09

**Product:** LDL-C Select FS  
**Product Code:** 14121

Cat. No.: 1 4121 99 10 021

---

**Delivery of kits:** No of kits: 6  
Kit Lot/Los: 60147472  
Kit Expiry: 2022-12-31

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 30723  
Expiry: 2022-12-31

Reagent 2 Lot No.: 30723  
Expiry: 2022-12-31

### Examination:

Appearance: Reagent 1: clear, yellowish to slightly pink  
Reagent 2: clear, yellowish

Odour: Reagent 1: no characteristic odour  
Reagent 2: no characteristic odour

**pH:** Reagent 1 ( 25 °C ± 1 °C): 6.76  
Reagent 2 ( 25 °C ± 1 °C): 6.97

**Absorbance:** Reagent 1: 0.002  
600-700 nm / 37 °C R1 + R2: 0.007

**Control sera:** TruLab L 1 TruLab L 2  
Lot: 29253 Lot: 29254

Test results 90.6 mg/dL 140 mg/dL

Target values 87.8 70.2 - 105 135 108 - 162

---

**Additional remarks:** Mf-Date: 02.07.2021

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka