CERTIFICADO DE ANÁLISE
CERTIFICATE OF ANALYSIS

MS ANVISA	**80115310042**
CÓDIGO / CODE	**107095**
COD. INTERNO / INTERNAL CODE	**1100075K**
DESCRIÇÃO / DESCRIPTION	**LACTATO R1 3X20ML + R2 1X15ML**

KIT

LOTE / LOT:	10709522145
DATA DE FABR./MANUF. DATE	07/2022
DATA DE VALIDADE / EXPIRY DATE:	08/2023

REAGENTE 1 / REAGENT 1

COD. INTERNO / INTERNAL CODE:	1100020R1K
LOTE / LOT:	10721322118
DATA DE VALIDADE / EXPIRY DATE:	08/2023

REAGENTE 2 / REAGENT 2

COD. INTERNO / INTERNAL CODE:	1100015R2K
LOTE / LOT:	10721422118
DATA DE VALIDADE / EXPIRY DATE:	08/2023

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	Conforme/According
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	Conforme/According
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	Conforme/According
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	Conforme/According
pH (25 oC) R1	8.90 - 9.10	Conforme/According
PH (25 oC) R2	2,50 - 2,70	Conforme/According
pH (25 oC) R1 + R2	8.75 - 9.15	Conforme/According
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	Conforme/According
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	Conforme/According
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	Conforme/According
TRULAB N - LOTE: 31253 (mg/dL)	8.55 - 13.1	11.90
TRULAB P - LOTE: 31254 (mg/dL)	25.1 - 38.5	31.53

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronically and is valid without a signature.

Monique Brasil Vianna de Sá
Controle de Qualidade / Quality Control