

CERTIFICADO DE ANÁLISE

CERTIFICATE OF ANALYSIS

MS ANVISA	80115310042
CÓDIGO / CODE	107320
COD. INTERNO / INTERNAL CODE	1100250K
DESCRIÇÃO / DESCRIPTION	LACTATO R1 5X40ML + R2 1X50ML
	KIT
LOTE / LOT:	10732023118
DATA DE FABR./MANUF. DATE	06/2023
DATA DE VALIDADE / EXPIRY DATE:	11/2024
	REAGENTE 1 / REAGENT 1
COD. INTERNO / INTERNAL CODE:	1100040R1K
LOTE / LOT:	10741623115
DATA DE VALIDADE / EXPIRY DATE:	11/2024
	REAGENTE 2 / REAGENT 2
COD. INTERNO / INTERNAL CODE:	1100050R2K
LOTE / LOT:	10741723115
DATA DE VALIDADE / EXPIRY DATE:	11/2024

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	CONFORME/ACCORDING
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	CONFORME/ACCORDING
pH (25 oC) R1	8.90 - 9.10	CONFORME/ACCORDING
PH (25 oC) R2	2,50 - 2,70	CONFORME/ACCORDING
pH (25 oC) R1 + R2	8.75 - 9.15	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	CONFORME/ACCORDING
TRULAB N - LOTE: 31253 (mg/dL)	9.17 - 14.0	11.5
TRULAB P - LOTE: 31254 (mg/dL)	24.1 - 36.9	31.1

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronically and is valid without a signature.

Andréa Araújo de Souza e Silva
Controle de Qualidade / Quality Control