

## Certificado de Análise Analysis Certificate

**MS ANVISA** 80115310205  
**DESCRIÇÃO / DESCRIPTION** CREATININA WS R1 2X40ML + R2 2X10ML  
**LOTE / LOT:** 112438239  
**DATA DE FABR./MANUF. DATE** 14/09/2023  
**DATA DE VALIDADE / EXPIRY DATE:** 09/03/2025

---

**DESCRIÇÃO / DESCRIPTION:** CREATININA WS R1 40ML  
**LOTE / LOT:** 112439239  
**DATA DE VALIDADE / EXPIRY DATE:** 23/08/2025

---

**DESCRIÇÃO / DESCRIPTION:** CREATININA WS R2 10ML  
**LOTE / LOT:** 112440239  
**DATA DE VALIDADE / EXPIRY DATE:** 09/03/2025

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
Aparencia / Appearance R1	LIQUIDO LIMPIDO INCOLOR/CLEAR, COLOURLESS LIQUID	Conforme/According
Aparencia / Appearance R2	LIQUIDO LIMPIDO AMARELADO/CLEAR, YELLOWISH LIQUID	Conforme/According
Odor / Odour R1	SEM ODOR CARACTERISTICO/NO CHARACTERISTIC ODOUR	Conforme/According
pH R1	13.0 (12.90-13.10)	13,0
Absorbancia / Absorbance R1	0.015 (0.000-0.030)	0,001
Absorbancia / Absorbance R1 + R2	0.150 (0.120-0.180)	0,152
Teste funcional / Functional test	DENTRO DOS VALORES DE REFERÊNCIA DOS CONTROLES/WITHIN THE REFERENCE VALUES OF THE CONTROLS	Conforme/According