

Certificado de Análise Analysis Certificate

MS ANVISA 80115310205
DESCRIÇÃO / DESCRIPTION CREATININA WS R1 4 X 38,6 ML + R2 4 X 11,4 ML (RESPONS)
LOTE / LOT: 1112192363
DATA DE FABR./MANUF. DATE 14/09/2023
DATA DE VALIDADE / EXPIRY DATE: 09/03/2025

DESCRIÇÃO / DESCRIPTION: CREATININA WS R1 38,6 ML + R2 11,4 ML
LOTE / LOT: 1112182363
DATA DE VALIDADE / EXPIRY DATE: 09/03/2025

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
Aparencia / Appearance R1	LIQUIDO LIMPIDO INCOLOR/CLEAR, COLOURLESS LIQUID	Conforme/According
Aparencia / Appearance R2	LIQUIDO LIMPIDO AMARELADO/CLEAR, YELLOWISH LIQUID	Conforme/According
Odor / Odour R1	SEM ODOR CARACTERISTICO/NO CHARACTERISTIC ODOUR	Conforme/According
pH R1	13.0 (12.90-13.10)	13,0
Absorbancia / Absorbance R1	0.015 (0.000-0.030)	0,001
Absorbancia / Absorbance R1 + R2	0.150 (0.120-0.180)	0,152
Teste funcional / Functional test	DENTRO DOS VALORES DE REFERÊNCIA DOS CONTROLES/WITHIN THE REFERENCE VALUES OF THE CONTROLS	Conforme/According