



**LORNE**  
LABORATORIES LIMITED

**Laboratórios Lorne Ltda.**

Unidade 1 Cutbush Park Industrial Estate  
Danehill, Lower Earley,  
Berkshire RG6 4UT, Reino Unido

Número do documento: F031

Nível de revisão: 01

Data: 31 AGO 2022

Título: CERTIFICADO DE ANÁLISE


Número da página: 1 de 1

## Certificado de Análise

<b>PRODUTO</b>	Reagente para Grupos Sanguíneos Anti-H Lectina
<b>NÚMERO DO LOTE</b>	115114-A1 e todos os sublotes (ou seja, 115114-A2, 115114-A3, 115114-A4, etc)
<b>DATA DE FABRICAÇÃO</b>	07/08/2024
<b>DATA DE VALIDADE</b>	07/08/2026
<b>CORANTE</b>	Nenhum
<b>CONDIÇÕES DE ARMAZENAMENTO</b>	Refrigerar a 2 – 8°C
<b>CARGA BIOLÓGICA</b>	Produto filtrado através de um filtro estéril de 0,2 µm. O reagente não é estéril
<b>CONSERVANTE</b>	< 0,1% de azida de sódio p/v

### DECLARAÇÃO DE CONTROLE DE QUALIDADE:

Este lote de Anti-H Lectina atende aos requisitos internos de especificidade.

<b>Revisor:</b>	<b>Assessora Científica</b>
<b>Assinatura:</b>	
<b>Data:</b>	<b>11 de outubro de 2024</b>



**LORNE**  
LABORATORIES LIMITED

**Lorne Laboratories Ltd.**

Unit 1 Cutbush Park Industrial Estate  
Danehill, Lower Earley,  
Berkshire RG6 4UT, UK

**Document Number: F031**

**Revision Level: 01**

**Date: 31 AUG 2022**

**Title: CERTIFICATE OF ANALYSIS**

**Page Number: 1 of 1**

## Certificate of Analysis

(See LOR006 for Methodology)

<b>PRODUCT</b>	Anti-H Lectin Blood Grouping Reagent
<b>LOT NUMBER</b>	115114-A1 and all sub-lots (i.e 115114-A2, 115114-A3, 115114-A4, etc)
<b>DATE OF MANUFACTURE</b>	2024-08-07
<b>DATE OF EXPIRY</b>	2026-08-07
<b>DYE</b>	None
<b>STORAGE CONDITIONS</b>	Refrigerate at 2 – 8°C
<b>BIOBURDEN</b>	Product filtered through a sterile 0.2 µm filter. The reagent is not sterile
<b>PRESERVATIVE</b>	< 0.1% Sodium Azide w/v

### QUALITY CONTROL STATEMENT:

This lot of Anti-H Lectin meets the in-house specificity requirements.

<b>Title</b>	<b>Lead Quality Assurance Coordinator</b>
<b>Signature</b>	Joyce Dos Santos Digitally signed by Joyce Dos Santos Date: 2024.09.27 14:11:46 +01'00'
<b>Date:</b>	<b>27 SEP 2024</b>

Printed copies are "uncontrolled"

It is the responsibility of the user to ensure that the correct version is used.  
Confidential and Proprietary Information of Lorne Laboratories LTD.