



Laboratórios Lorne Ltda.

Unidade 1 Cutbush Park Industrial Estate
Danehill, Lower Earley,
Berkshire RG6 4UT, Reino Unido

Número do documento: F031

Nível de revisão: 01

Data: 31 AGO 2022

Título: CERTIFICADO DE ANÁLISE

Número da página: 1 de 1

Certificado de Análise

PRODUTO	Reagente Anti-A1 Lectina
NÚMERO DO LOTE	116147-B5 (resultados obtidos a partir do lote mãe 116147-B1)
DATA DE FABRICAÇÃO	04/11/2025
DATA DE VALIDADE	04/11/2027
CORANTE	Nenhum
CONDIÇÕES DE ARMAZENAMENTO	Refrigerar a 2 – 8°C
CARGA BIOLÓGICA	Produto filtrado através de um filtro estéril de 0,2 µm. O reagente não é estéril
CONSERVANTE	< 0,1% de azida de sódio p/v

CONTROLE DE QUALIDADE:

POTÊNCIA:	Células A ₁		Título	
			1 em 8	
ESPECIFICIDADE:	Fenótipos positivos		Fenótipos negativos	
	Células A ₁	Grau 5	Células A ₂	Negativo
	Células A ₁ B	Grau 5	Células A ₂ B	Negativo
			Células B	Negativo
			Células O	Negativo

DECLARAÇÃO DE CONTROLE DE QUALIDADE

Este lote de Anti-A1 atende aos requisitos de potência e especificidade descritos na edição atual das Diretrizes para os Serviços de Transfusão de Sangue no Reino Unido.

Revisor:	Coordenadora de Assessoria Científica
Assinatura:	<i>Alexandra Pereira</i>
Data:	25 de fevereiro de 2026



LORNE
LABORATORIES LIMITED

Lorne Laboratories Ltd.

Unit 1 Cutbush Park Industrial Estate
Danehill, Lower Earley,
Berkshire RG6 4UT, UK

Document Number: F031

Revision Level: 01

Date: 31 AUG 2022

Title: CERTIFICATE OF ANALYSIS

Page Number: 1 of 1

Certificate of Analysis

(See LOR006 for Methodology)

PRODUCT	Anti-A1 Lectin
LOT NUMBER	116147-B5 (Results are obtained from the mother lot 116147-B1)
DATE OF MANUFACTURE	2025-11-04
DATE OF EXPIRY	2027-11-04
DYE	None
STORAGE CONDITIONS	Refrigerate at 2 – 8°C
BIOBURDEN	Product filtered through a sterile 0.2 µm filter. The reagent is not sterile
PRESERVATIVE	< 0.1% Sodium Azide w/v

QUALITY CONTROL:

POTENCY:	A₁ Cells	Titer
		1 in 8
SPECIFICITY:	Positive Phenotypes	Negative Phenotypes
	A₁ Cells Grade 5	A₂ Cells Negative
	A₁B Cells Grade 5	A₂B Cells Negative
		B Cells Negative
		O Cells Negative

QUALITY CONTROL STATEMENT

This lot of Anti-A₁ meets the potency and specificity requirements as described in the current edition of the Guidelines for the Blood Transfusion Services in the UK.

Title	Quality Assurance Coordinator	
Signature	sangeetha natarajan	Digitally signed by sangeetha natarajan Date: 2026.01.22 10:05:28 Z
Date:	22 JAN 2026	