



**Lorne Laboratories Ltd.**  
Unidade 1 Parque Industrial Cutbush Park  
Danehill, Lower Earley,  
Berkshire RG6 4UT, Reino Unido

<b>Número do Documento: F031</b>	<b>Nível de Revisão: 01</b>	<b>Data: 31 AGO 2022</b>
<b>Título: CERTIFICADO DE ANÁLISE</b>		<b>Número da página: 1 de 2</b>

## Certificado de Análise

<b>PRODUTO</b>	Controle D Negativo Monoclonal
<b>NÚMERO DO LOTE</b>	650421-A3
<b>DATA DE FABRICAÇÃO</b>	15/01/2026
<b>DATA DE VALIDADE</b>	15/07/2028
<b>CORANTE</b>	Nenhum
<b>CONDIÇÕES DE ARMAZENAMENTO</b>	Refrigerar de 2 à 8°C
<b>CARGA BIOLÓGICA</b>	Produto filtrado por um filtro estéril de 0,2 µm. Não fornecido estéril.
<b>CONSERVANTE</b>	<0,1% de azida de sódio p/v

### CONTROLE DE QUALIDADE:

<b>ESPECIFICIDADE: Rhesus Fenótipos</b>	<b>Tubo Teste</b>	<b>Cartão BioVue</b>	<b>Cartão Diamed</b>
R1R1	Negativo	Negativo	Negativo
R2R2	Negativo	Negativo	Negativo
R1R2	Negativo	Negativo	Negativo
Du	Negativo	Negativo	Negativo
Dvi	Negativo	Negativo	Negativo

<b>DECLARAÇÃO DE CONTROLE DE QUALIDADE</b>	Esse lote de Controle D Negativo Monoclonal atende a todas as especificações internas para características sorológicas e físicas.
--	---

Revisor:	Coordenadora de Assessoria Científica
Assinatura:	<i>Alessandra Pereira</i>
Data:	06 de março de 2026



**LORNE**  
LABORATORIES LIMITED

**Lorne Laboratories Ltd.**  
Unit 1 Cutbush Park Industrial Estate  
Danehill, Lower Earley,  
Berkshire RG6 4UT, UK

**Document Number: F031**

**Revision Level: 01**

**Date: 31 AUG 2022**

**Title: CERTIFICATE OF ANALYSIS**

**Page Number: 2 of 2**

## Certificate of Analysis

(See LOR006 for Methodology)


<b>PRODUCT</b>	Monoclonal D Negative Control
<b>LOT NUMBER</b>	650421-A3
<b>DATE OF MANUFACTURE</b>	2026-01-15
<b>DATE OF EXPIRY</b>	2028-07-15
<b>DYE</b>	None
<b>STORAGE CONDITIONS</b>	Refrigerate at 2 – 8°C
<b>BIOBURDEN</b>	Product filtered through a sterile 0.2 µm filter. Not supplied sterile.
<b>PRESERVATIVE</b>	<0.1% Sodium Azide w/v

### QUALITY CONTROL:

<b>SPECIFICITY:</b>	<b>Rhesus Phenotypes</b>	<b>Tube Test</b>	<b>BioVue Card</b>	<b>Diamed Card</b>
	R <sub>1</sub> R <sub>1</sub>	Negative	Negative	Negative
	R <sub>2</sub> R <sub>2</sub>	Negative	Negative	Negative
	R <sub>1</sub> R <sub>2</sub>	Negative	Negative	Negative
	D <sup>u</sup>	Negative	Negative	Negative
	D <sup>vi</sup>	Negative	Negative	Negative

### QUALITY CONTROL STATEMENT

This lot of Monoclonal D Negative Control meets all in-house specifications for serological and physical characteristics.

<b>Title</b>	Quality Assurance Coordinator
<b>Signature</b>	<b>sangeetha natarajan</b>  Digitally signed by sangeetha natarajan Date: 2026.01.23 09:24:30 Z
<b>Date:</b>	23 JAN 2026